

GOVERNMENT OF ASSAM
OFFICE OF THE PRINCIPAL CUM CHIEF SUPERINTENDENT
JORHAT MEDICAL COLLEGE & HOSPITAL, JORHAT
(UNDER SOCIETY FOR MEDICAL EDUCATION, JORHAT)

No. SMEJ / JMCH / 1180/Pt.1/2014/2896

Dated Jorhat, the 15th October 2016

SHORT QUOTATION NOTICE

Sealed Quotations affixing court fee stamps of Rs. 8.25 (Rupees Eight and Paise Twenty Five) only are invited from manufacturers or their accredited dealers for supply and Installation of CCTV and Binocular Biological Microscope(student) as per Annexure-I & II to be procured for Jorhat Medical College & Hospital, Jorhat.

The Quotation(s) are to be submitted / dropped in the box kept in the office of the undersigned on or before 01:00 PM of 29.10.2016 and shall be opened on the same day at 02:00 PM in presence of the quotationers or their authorized representatives. The Quotationers are also required to submit a non-refundable amount of Rs.1,000/- (Rupees One Thousand) only by way of Demand Draft to be drawn in favour of Jorhat Medical College Hospital Users' Fund, Jorhat as 'Cost of Document'. Details of quotations shall be available and can be downloaded from the official website : jmch.org.in from 20.10.2016.


Documents to be submitted along with Quotations:

1. Demand Draft for Rs.1,000/- (Rupees One Thousand) only to be drawn in favour of Jorhat Medical College Hospital Users' Fund, Jorhat.
2. Sales Tax Clearance Certificates up to 31st March 2016.
3. Financial ability Certificate from competent authority along with minimum annual turnover of Rs. 01.00 crore (Rupees one crore only).
4. Tenderer's must be having prior working experience (not less than three years) in Govt / Private organization. They have to submit performance certificate from that particular organization.
5. Registration Certificate of the Firm under Society Act; in case of Society.
6. Photocopy of PAN Card / GRN No. / TIN No.
7. A sum of Rs. 5,000/- (Rupees Five Thousand) only shall have to be submitted by way of Demand Draft favouring 'Jorhat Medical College Hospital User Fund, Jorhat' as Earnest Money along with the Quotation.
8. The tenderer must have executed similar work before should have a minimum order of Rs. 40.00 lakhs in a single order in the last three years.

Terms & Conditions of the Quotations :

1. Prices should be quoted item-wise in each group separately. Care should be taken to avoid mix-up of items as shown in the group.
2. The Bidder must strictly follow the specifications as provided for supplying the items.
3. Prices should be quoted as per item as stated in Annexure-I&II.
4. Prices should be shown both in figure and words neatly and legibly. Avoid overwriting.
5. Taxes, as applicable, must be shown separately.
6. Only the lowest quoted quotation will be accepted and supply orders will be given accordingly.
7. The undersigned reserves the right to accepts or reject any or all the quotations without showing any reasons thereof.
8. The rates approved shall be valid for 2 (two) year.
9. Compulsory training, if required, shall have to be provided after complete installation of the item/ machines for minimum a week's time.
10. No refurbished item or software will be acceptable. A declaration is to be made by the bidder that they shall not supply any refurbished item or software.
11. The company / firm should provide regular software updates, whenever released, free of cost for the entire life of the item / machine.
12. The bidder should furnish necessary certificate of authorised dealership / distributorship etc.
13. The company / firm should indicate clearly necessary terms of warranty of the supplied machines which will be for at least 2 years from the date of installation and 3 years CMC after the expiry of the warranty.


Principal cum Chief Superintendent
Jorhat Medical College & Hospital.


15/10/2016

GOVERNMENT OF ASSAM
OFFICE OF THE PRINCIPAL CUM CHIEF SUPERINTENDENT
JORHAT MEDICAL COLLEGE & HOSPITAL, JORHAT
(UNDER SOCIETY FOR MEDICAL EDUCATION, JORHAT)


ANNEXURE : I
Technical Bid/Specification of CCTV

	<u>Technical Bid/ Specification Of CCTV</u>
<u>A</u>	<u>Cameras</u>
1	Night Vision Dome HD 20Mtr
2	Night Vision Dome HD 30Mtr
3	Night Vision Bullet HD 50Mtr
4	Night Vision Bullet HD 80Mtr
5	Outdoor Speed Dome Camera Rotatable 12X Zoom
6	Outdoor Night Vision Speed Dome Camera Rotatable 20X Zoom'
<u>B</u>	<u>DVR (Recorder)</u>
7	4Ch Stand Alone Digital Video Recorder
8	8Ch Stand Alone Digital Video Recorder
9	16Ch Stand Alone Digital Video Recorder
<u>C</u>	<u>Power Supply</u>
10	12Volt 2AMP
11	12Volt 5AMP
12	12Volt 10AMP
<u>D</u>	<u>Display</u>
13	15.6"
14	18.5"
15	21.5"
16	32"
17	50"
<u>E</u>	<u>Wire And Accessories</u>
18	3+1
19	RJ59 + 2Core
20	BNC Connector
21	Power Connector
22	2U Rack
23	Iron Pole 20 Feet
24	UPS 1KVA
25	Casing Bit

B
15/11/2016

GOVERNMENT OF ASSAM
OFFICE OF THE PRINCIPAL CUM CHIEF SUPERINTENDENT
JORHAT MEDICAL COLLEGE & HOSPITAL, JORHAT
(UNDER SOCIETY FOR MEDICAL EDUCATION, JORHAT)

F	<u>Storage</u>
26	1TB HDD
27	2TB HDD
28	3TB HDD
29	4TB HDD
E	<u>Optional</u>
30	Optical Fiber Cable
31	PLB
32	Media Converter
33	LAN Gigabyte Switch
34	OFC Patch Cable
35	LAN Patch Cable
36	Server (Xeon Processor/ 4GBRAM)
37	HDMI Cable
38	Rack


Principal cum Chief Superintendent
Jorhat Medical College & Hospital.


১১/১০/২০১৬

GOVERNMENT OF ASSAM
OFFICE OF THE PRINCIPAL CUM CHIEF SUPERINTENDENT
JORHAT MEDICAL COLLEGE & HOSPITAL, JORHAT
(UNDER SOCIETY FOR MEDICAL EDUCATION, JORHAT)

Annexure: II

Technical Bid of Binocular Biological Microscope (Student)

1. Observation Method : Brightfield/Darkfield
2. Illuminator : Transmitted Koehler Illuminator, Lamp: 30 W
3. Focus : Focusing Mechanism/Stage Focus
Coarse Handle Stroke: 25 mm
Coarse Handle Stroke per rotation: 36.8 mm
4. Features : Stage height movement by roller guide (rack & pinion)
Upper limit stopper
Tension adjustment on coarse focus adjustment knob.
5. Revolving : Manual/Standard Type: Built-in 4 Position.
6. Stage : Mechanical/Mechanical stages with Right-Hand control
:Built in: X: 76 mm, Y: 50 mm
7. Condenser : Manual /Abbe Conenser: NA 1.25/W.D.-(4X-100X)(Built-in)
8. Observation Tubes : Widefield(FN 22)/Binocular/Tilting Binocular.
9. Tube Including Angel : 30°
10. Trinocular Tube light (Optional): -
Path selection (camera observation)
11. Interpupillary Distance : 48-75 mm
Adjustmnt
12. Dimensions weight : 233 (W) x 367.5 (D) x 411(H)mm, 8 Kg (standard
configuration)



Prof. (Dr.) D. Hazarika
Principal cum Chief Superintendent
Jorhat Medical College & Hospital, Jorhat

১০/১০/২০১৬